

## Highlights

- Detection of non-electroactive melamine using a Au-Ag/Nafion/GCE
- Uric acid anodic response is reduced by melamine/Au–Ag NPs hydrophobic interaction
- The detection limit for melamine was as low as 1.8 nM
- The proposed method could detect melamine in milk powder samples with good recovery